

# COVID-19 Antigen Rapid Test Cassette (Nasal Swab) Clinical Sensitivity and Specificity Study Report

## 1. Objective

The CLUNGENE<sup>®</sup> COVID-19 Antigen Rapid Test Cassette (Nasal Swab) (hereinafter referred to as the CLUNGENE Device) manufactured by Hangzhou Clongene Biotech Co., Ltd. is a lateral flow immunoassay intended for the qualitative detection SARS-CoV-2 nucleocapsid antigens in nasal swab from individuals who are suspected of COVID-19 by their healthcare provider.

This study is intended to evaluate the clinical performance, between the CLUNGENE Device and the comparator RT-PCR assay.

## 2. Method

A study of 617 direct nasal swabs was performed. The specimens were prospectively collected from patients in community meeting Department of Public Health definition of a suspected COVID-19 case and being tested for SARS-CoV-2 part of routine medical care at 5 locations and tested at a single central laboratory.

Two nasal swabs were collected from individual symptomatic (within 7 days of onset) or asymptomatic patients who were suspected of COVID-19. At all locations, one nasal swab was tested directly with the COVID-19 Antigen Rapid Test Cassette (Nasal Swab) according to product instructions for use, and the other swab was eluted in 3 mL viral transport media and tested with RT-PCR assay for detection of SARS-CoV-2. Swabs were randomly assigned to testing with the COVID-19 Antigen Rapid Test Cassette (Nasal Swab) or RT-PCR assay and were tested by operators who were blinded to the RT-PCR test result.

The positive percent agreement (PPA) was calculated as  $100\% \times (\text{True Positive} / [\text{True Positive} + \text{False Negative}])$ . The negative percent agreement (NPA) was calculated as  $100\% \times (\text{True Negative} / [\text{True Negative} + \text{False Positive}])$ . Accuracy was calculated as  $100\% \times ([\text{True Positive} + \text{True Negative}] / \text{Total sample Qty})$ . The 95% (two-sided) confidence interval (CI) was calculated using the Wilson Score Method.

## 3. Comparator method

Real-Time Fluorescent RT-PCR Kit for Detecting SARS-CoV-2, manufactured by BGI Genomics Co. Ltd., is a real-time reverse transcription polymerase chain reaction (rRT-PCR) test. This product has got CE, NMPA certification and FDA Emergency Use Authorized. A specimen is positive for SARS-CoV-2 if the Ct value of ORF1ab gene is not higher than 37 and the Ct value of human housekeeping gene  $\beta$ -Actin is not higher than 35.

## 4. Enrollment criteria (inclusion/exclusion criteria)

### 4.1 Inclusion criteria

- Patients in community meeting Department of Public Health definition of a suspected COVID-19 case and being tested for SARS-CoV-2 part of routine medical care.
- Symptomatic (within 7 days of onset) or asymptomatic patients who were suspected of COVID-19.

### 4.2 Exclusion criteria

- Unable to obtain samples of information needed for the experiment
- Samples that have been contaminated or contaminated during sample storage
- Samples with inappropriate storage conditions

## 5. Result

The results are summarized in the following table.

The RT-PCR cycle threshold (Ct) is the relevant signal value. Lower Ct value indicate higher viral load. The sensitivity was calculated for the different Ct value range (Ct value  $\leq 30$  and Ct value  $\leq 37$ ).

COVID-19 Antigen		RT-PCR (Ct value $\leq$ 30)		Total
		Positive	Negative	
<b>CLUNGENE<sup>®</sup></b>	Positive	117	3	120
	Negative	3	462	465
Total		120	465	585

PPA (Ct $\leq$ 30):97.5% (117/120), (95%CI: 92.9%~99.2%)

NPA: 99.4% (462/465), (95%CI: 98.1%~99.8%)

COVID-19 Antigen		RT-PCR (Ct value $\leq$ 37)		Total
		Positive	Negative	
<b>CLUNGENE<sup>®</sup></b>	Positive	139	3	142
	Negative	13	462	475
Total		152	465	617

PPA (Ct $\leq$ 37):91.4% (139/152), (95%CI: 85.9%~94.9%)

NPA: 99.4% (462/465), (95%CI: 98.1%~99.8%)

PPA - Positive Percent Agreement (Sensitivity)

NPA - Negative Percent Agreement (Specificity)

## 6. Conclusion

Taken together, the CLUNGENE<sup>®</sup> COVID-19 Antigen Rapid Test Cassette (Nasal Swab) had a positive percent agreement (sensitivity) of 97.5% (95%CI: 92.9%~99.2%) with specimens of a Ct count  $\leq$ 30, 91.4% (95%CI: 85.9%~94.9%) with specimens of a Ct count  $\leq$ 37, negative percent agreement (specificity) of 99.4% (95%CI: 98.1%~99.8%).